


# Finisterre

by Teknokroma™

## Certificate of Analysis and Conformance SPE Cartridges

Material	C18 500 mg/ 6ml. 267880
Phase Type	Octadecylsilyl
Endcapped	Yes

Test	Specification	Result
Average Particle Size	30-70 $\mu\text{m}$	52 $\mu\text{m}$
Average Pore Size	53-80Å	65 Å
Silica surface area	510-580 m <sup>2</sup> /g	563 m <sup>2</sup> /g
Elemental analysis	17-18 %C	17,49 %C
Extract pH (dH <sub>2</sub> O)	4-7	5,4



This product complies all procedures, manufacturing specifications and quality controls for release.

Signed

Quality Manager



Date

November 25, 2020

Product manufactured in compliance with ISO 9001 and technical procedures.

Not for human use. Liability limited to the restitution of the product.

